NEVO+600ML LOW NOISE MEDICAL DATA SHEET

AC/DC Modular Configurable PSU





600 Watts in the palm of your hand

The NEVO+600ML modular configurable medical power supply is the smallest in its class and the ultimate solution for demanding medical applications where size, power density and weight matter. Its tiny footprint of 5" x 3" x 1.61" weighs only 600 grams and delivers an incredible 450 Watts with a minimum of audible noise. The input module can accommodate up to four isolated output modules which can be configured into a high power 5"x 3" single output power supply or a multiple output power supply with up to 8 isolated outputs. Standard features include intelligent fan control providing optimised airflow for various load and temperature conditions, wide output voltage adjust, parallel and series connection of modules and an isolated 5V 1A bias supply. The low noise fan option allows you to use this innovative power supply in even the quietest of environments. The series is approved to latest medical standards and features market leading specifications and design in application support.

MAIN FEATURES

• 450 Watts in 5" x 3" x 1.61"	 Up to 8 isolated outputs 	 Accurate current sharing
 Low noise operation (~18dBA reduction from S version) 	 User and field configurable 	 Standard 5V 1A bias supply
Intelligent fan control	 Parallel and series connection of modules 	 IEC/UL60601-1 Ed. 3 & -1-2 Ed. 4 (EMC)
• Efficiency up to 89%	 Wide output voltage adjust range 	• 3 Year warranty
	 Remote current / voltage programming 	 Parallel units with droop current sharing
APPLICATIONS		
Medical & diagnostic equipment	 Telecommunications 	Lasers
 Test & Measurement equipment 	 Laboratory & Analysis equipment 	LED lighting
Robotics	• Display	 Retrofit of legacy PSUs
• Oil & Gas	Avionics	
CUSTOMER BENEFITS		
Fast time to market	 Proven technology 	 Technology consolidation
 24 hrs samples from distribution 	Eliminates custom design costs	Supplier consolidation
 Safety & EMC certified 	 Field replaceable 	
 World class engineering support 		

Page 1 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

SPECIFICATIONS

INPUT MODULE SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
AC Input Voltage	Nominal range is 100V _{RMS} to 240V _{RMS}	85		264	V _{RMS}	
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz	
DC Input Voltage	Not covered by safety approvals. Contact Vox Power.	120		300	V _{DC}	
Output Power Rating	De-rate linearly from 450Watts at 120V _{RMS} to 338Watts at 85V _{RMS}			450	Watts	
Input Current	450Watts output at 120 V _{RMS} input			5	Amps	
Input Current Limit	Maintains power factor		8		Amps	
Inrush Current	265V _{RMS} , 25°C (cold start)			20	Amps	
Fusing	Live line fused (5x20 Fast acting)			8	Amps	
Efficiency	See graphs		86	89	%	
No load Power consumption	All outputs fitted and disabled/enabled		21/28		Watts	
Power Factor	Typical value for 300 Watts output at 240Vrms input		0.96	0.99		
Holdup	450Watts output at 120V _{RMS} input	17	20	21	mS	
UVP	Turn on under voltage protection	78		84	V _{RMS}	
Over temperature	Internally monitored.	115		125	°C	
Reliability (1)	Input module			1.207	FPMH	
	Fan			2.7	FPMH	
Warranty	Standard terms and conditions apply			3	Years	
Size	133.7 (L) x 77.7 (W) x 41.0 (H). See diagram for tolerance details				mm	
Weight	360 + 60 per output module				Grams	
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Control	led				

GLOBAL SIGNALS SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
Bias Voltage	Two isolated Bias Outputs available	4.8	5	5.2	Volts	
Bias Current	Hiccup type current limit	0		1	Amps	
AC_OK Voltage	Low output level High output level	0 3.5	0.2 4.5	1 5.2	Volts	
AC_OK Current		-10		20	mA	
Power Good Voltage	Low output level. internal 10kΩ pull down. High output level. PNP open collector.	0 8	0 10	0 15	Volts	
Power Good Current	Open collector output. Current source only. All Slots.			20	mA	
Global Inhibit Voltage	Low input level High input level	0 3		1 15	Volts	
Global Inhibit Current	5k input impedance.	0.6		3	mA	
Inhibit Voltage	Low input level. All slots. High input level. All slots.	0 2.5		1 15	Volts	
Inhibit Current	10k input impedance. All slots.	0.25		1.5	mA	

OUTPUT MODULE SPECIFICATION SUMMARY												
MODEL ·	Out	put Volta	age	Output	Rated	Peak	Load	Line	Cross	Ripple &	FPMH ⁽¹⁾	Feature
MODLL	Min.	Nom.	Max.	Current	Power	Power	Reg.	Reg.	Reg.	Noise		Set (2)
OP1	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV _{PP}	0.5	ABCDEFG
OP2	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFG
OP3	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	240mV _{PP}	0.5	ABCDEFG
OP4	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV _{PP}	0.5	ABCDEFG
OP5	3.3V	12V	15V	5A	2x 75W	2x 75W	±50mV	±12mV	±24mV	240mV _{PP}	0.75	AFG
OP8	23.2V	24V	24.7V	3.125A	2x 75W	2x 75W	±100mV	±24mV	±48mV	480mV _{PP}	0.75	AFG
OPA2 ⁽³⁾	4.5V	12V	15V	25A	300W	375W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFGH
OPA3(3)	9V	24V	30V	15A	300W	450W	±150mV	±24mV	±48mV	$240 mV_{PP}$	0.5	ABCDEFGH
Note 1.	Note 1. Output module, 30°C base, 100% load, SR332 issue 2 Method I, Case 3, Ground, Fixed, Controlled											
Note 2.	A = Rem	ote Sense, l	B = Extern	al Voltage contro	ol, C = External	constant curre	ent control, D	= Current ou	ıtput signal, E	= Current share,	F = Over Voltag	e protection,

G = Over temperature protection, H = Dual Slot module

Note 3. Can only be used with NEVO+600 chassis with date codes from 2048 onwards. eg. 2048C080000 can use A2 or A3 module, 2047C089999 cannot use A2 or A3 module.

SAFETY SPECIFICATIONS					
Parameter	Details	Max	Units		
	Input to Output (2 MOPP). Do not perform test on assembled unit ⁽¹⁾	4000	V _{AC}		
Isolation Voltages	Input to Chassis (1 MOPP)	1500	V _{AC}		
	Global signals (J2) to Output/Chassis	250	V _{DC}		
	Output to Output/Chassis (Standard modules)	250	V _{DC}		
Earth Leakage Current	Normal condition, 264Vac, 63Hz, 25°C	300	uA		
Touch Leakage Current	Standard modules NC/SFC	20/200	uA		
Patient Leakage Current	Standard modules 264Vac, 63Hz, 25°C NC/SFC ⁽²⁾		uA		
Note 1. Testing an assembled unit to 4000V _{AC} may cause damage. Please refer to application note (APN-002) on Vox Power website or contact Vox Power representative.					
Note 2. Not Applicable					


INSTALLATION SPECIFICATIONS						
Parameter Details Parameter Details						
Equipment class		Flammability Rating	94V-2			
Overvoltage category	II	Ingress protection rating	IP10			
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU & 2015/863/EU			
Pollution degree	2	Intended usage environment	Home Healthcare			


Page 2 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com

		ENVIRONMENTAL SPE	CIFICATION	S					
Developmenter	Detaile		No	Non-Operational			Operational		
Parameter	Details	Details -			Min Max		Max	Units	
Air Temperature	Operational limits subject to	appropriate de-ratings	-4		+85	-20	70	°C	
Humidity	Relative, non-condensing		5		95	5	95	%	
Altitude			-20		5000	-200	3000	m	
Air Pressure			5.	2	106	69	106	kPa	
Noise Level	Variable. Measured 1m from		-		-	18	42	dBA	
Shock	3000 bumps at 10G (16ms) half sine wave								
Vibration	1.5G 10 to 200Hz sine wave,	20G for 15min in 3 axes random vibration							
		ELECTROMAGNETIC COMPLI	ANCE – EMI	SSIOI	٩S				
Phenomenon		Basic EMC Standard		Test	Details				
Radiated emission	ns, electric field	EN55011/22, FCC		Class	B compliant				
Conducted emissi	ons	EN55011/22, FCC part 15, CISPR 2	2/11	Class	B compliant				
Harmonic Distorti	on	IEC61000-3-2	Compliant						
Flicker & Fluctuati	on	IEC61000-3-3		Compliant					
		ELECTROMAGNETIC COMPL	IANCE – IMN	NUNI	ΓY				
Phenomenon		Basic EMC Standard		Test	Details				
Electrostatic disch	arge	IEC61000-4-2	Test level 4	Test level 4: 15kV air, 8kV contact					
Radiated RF EM fie	elds	IEC61000-4-3	Test Level	Test Level 3: (10V/m, 80MHz-2.7GHz) sine wave AM 80% 1kHz					
Proximity fields fro equipment	om RF wireless communications	IEC61000-4-3	Test levels	Test levels as per IEC60601-1-2:2014 Table 9					
Electrical Fast Tran	nsients/bursts	IEC61000-4-4	Test Level	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4)					
Surges		IEC61000-4-5	Test Level	Test Level 3: 1kV L-N, 2kV L-E					
Conducted distur	bances induced by RF fields	IEC61000-4-6	Test Level	Test Level 3: 10V, 0.15 to 80Mhz sine wave AM 80% 1kHz					
Power Frequency	Magnetic Fields	IEC61000-4-8	Test level 4	4: 30A/n	n 50Hz				
Voltage Dips		IEC61000-4-11& SEMI-F47-0706 ⁽²		0% 10ms, 0% 20ms, 80% 1s, 80% 10s, 90% continuous (Criterion A) 70% 0.5s, 40% 0.2s (Criterion A at 240V and Criterion B at 100V)					
Voltage interrupti	ons	IEC61000-4-11	0% 250/30	0 cycle	as per IEC60	601-1-2:2014	(Criterion B)		
Notes: 1.	Criterion B = Temporary degr Criterion C = Temporary loss	of performance or loss of function. adation of performance or loss of function is of function is allowed but requires operator i V to 240V). Line deratings applied where app	ntervention to re		unction is sel	f-recoverable	ŀ.		
		AGENCY APPR	·						
Standard		Details					File		
IEC 60601-1:2005 + A1:2012	+ CORR1 2006 + CORR2: 2007	Medical electrical equipment Part 1: Gener performance	al requirements	for basi	c safety and	essential	UL: E31	6486	

+ A1:2012	performance	
EN60601-1:2006 + A11:2011 + A1:2013 +	Medical electrical equipment Part 1: General requirements for basic safety and essential	
A12:2014	performance	
CAN/CSA-C22.2 No. 60601-1 (2008)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential	
CAN/CSA-C22.2 NO. 00001-1 (2008)	Performance	
ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential	
ANSI/AAMI ES00001-1 (2005 + C1.09 + A2.10)	Performance	
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU	


CB certificate and report available on request





Page 3 of 5 Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com


MECHANICAL DIMENSIONS AND MOUNTING SCREWS





CONNECTORS

PIN	

PINOUTS						
J1						
Circuit	Details					
1	Live					
2	Earth					
3	Neutral					
	J2					
Circuit	Details					
1	Power good	Slot A				
2	Inhibit	SIOLA				
3	Power good	Slot B				
4	Inhibit	3101.0				
5	Power good	Slot C				
6	Inhibit	5101 C				
7	Power good	Slot D				
8	Inhibit	510LD				
9	Global inhibit					
10	AC OK					
11	+5V 1A bias supply					
12	COM					
	J5 ⁽⁴⁾					
Circuit	Details					
1	-Sense					
2	+Sense					
3	Voltage control					











Positive output

REF.	DETAILS	MANUFACTURER	HOUSING	TERMINAL				
J1	MAINS INPUT: 3 Pin, 5.08mm, with Friction Lock, 18-24 AWG	MOLEX	10013036	0008701031				
J2	GLOBAL SIGNALS: 12 Pin, 2mm, without Friction Lock, 24-30 AWG	MOLEX	511101251	0503948051				
J3/4 ⁽¹⁾	OUTPUT POWER TERMINAL: TAB SIZE 6.35mmx0.8mm	VARIOUS		VARIOUS				
J5	OUTPUT SIGNALS: 6 Pin, 1.25mm, with Friction lock, 28-32 AWG	MOLEX	0510210600	0500588000				
Notes								
1. Termin	1. Terminal and wire current rating must exceed maximum short circuit output current. Eg. Output 1 = 25A*1.25 = 31.25Amps							
2. Direct e	equivalents may be used for any connector parts							

3. All cables must be rated 105°C min, equivalent to UL1015

Current control / share / out


+5V 10mA local bias supply

5

COM

4. Pinout is for single output types only

PART NUMBERING SYSTEM



All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLAIMS ALL WARRANTIES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTIES OF SUITABILITY, FITNESS FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT.

Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or military or naval missile or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.

Vox Power Limited | Unit 2, Red Cow Interchange Estate, Ballymount, Dublin 22, D22 Y8H2, Ireland | T +353 1 4591161 | www.vox-power.com Page 5 of 5